



California Medical Device Recall Information



Recall Name

CLR Medicals International Recalls Viscoelastic Products Due to Safety Risks

Recall Date	Product Description	Recalling Firm	Recall Reason
11/25/13	<ul style="list-style-type: none">• Viscocel• Viscocel Plus	CLR Medicals International, Inc. Pomona, CA	<i>Violations of good manufacturing practices related to sterility controls, and lack of FDA approval of products as safe and effective for distribution within the United States.</i>
Recall Class	Product Identification	Distribution	Affected Dates
TBD	ALL products within the labeled shelf-life are recalled; Expired product should be discarded.	Undetermined	Manufactured from: September 1, 2011 to September 19, 2013. Distributed: October 2011 to November 2013.

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/Safety/Recalls/ucm376736.htm>